



Replacement Oxygen Sensors



Cross Reference List

Advanced Instruments Inc.

Manufacturer	Part No.	Response Time in Seconds	Warranty / Expected Life		Replacement Part No.	Response Time in Seconds	Warranty / Expected Life
Teledyne ppm	A-2C	7	6 / 8		XLT-12-123	7	12 / 24
	B-2 (a)	(a)	(a)		PSR-12-22	7	12 / 24
	B-2C	7	6 / 8		PSR-12-223	7	12 / 24
	B-4	7	6 / 8		PSR-12-24	7	12 / 24
	L-2	7	12 / 18		(b)	7	12 / 24
	S-2 / Z-2	5-10	6 / 8		(c)	(c)	(c)
Teledyne %	A-3 / A-5	45	6 / 8		XLT-11-15	30	12 / 15
	B-1	7	6 / 8		PSR-11-21	7	9 / 12
	B-1C	7	6 / 8		PSR-11-213	7	9 / 12
	B-3	13	12 / 12		PSR-11-23	13	12 / 15
	B-3C	13	12 / 12		PSR-11-233	13	12 / 15
	B-5	13	18 / 36		PSR-11-25	10	24 / 48
	B-7	7	12 / 24		(d)	(d)	(d)
	B-9	10	18 / 36		PSR-11-77	6	24 / 42
	C-3	30	12 / 18		PSR-11-54	7	18 / 24
	C-5	30	18 / 45		(e)	(e)	(e)
	E-2	20	24 / 48		PSR-11-52	10	24 / 60
MSA % Sensors	466324	13	12 / 15		PSR-11-39	6	24 / 32
	480566	13	12 / 15		PSR-11-37-4	9	24 / 32
Figaro % Sensor	KE-25	12	12 / 60		PSR-11-75-KE	9	24 / 60

Response Time: 90% of full scale
 Warranty: Competitor's pro-rated
 Expected Life: In air at 25°C, 1 atm

(a) Discontinued, see B2C
 (b) Use PSR-12-223 + A-2363 spacer
 (c) Use PSR-12-223

(d) Use PSR-11-54
 (e) Use PSR-11-25
 * Subject to change without notice